



The Testing and Calibration Laboratory
of TÜV SÜD Industrie Service GmbH

confirms to the company

Disruptive Technologies Research AS
Sandakerveien 17
1366 Lysaker, Norway

due to the positive results of the tested temperature recorder,

Cloud Connector with wireless Sensor

according to the test-report No. **K-P-TS-26-200** dated 2026-06-12, that the
temperature recorder fulfils the requirements of the
standard DIN EN 12830:2018-10.

Sensor type	Wireless Sensor EU
Accuracy class	0,5
Measurement range	-30°C to +50 °C
Application range	Storage
Software verification level	IIIb

The manufacturer is allowed to mark the temperature recorder with the following
test-report number:

ATP-MUC K-P-TS-26-200

This certificate is valid till: 2031-06-30

(As long as the normative requirements, the declared fields of application, the product and the production conditions are not significantly changed)

Munich, the 2026-06-12

Center of Competence
for Refrigeration and Air-conditioning
Head of test laboratory

Heiko Mirring

Expert

Thorsten Bennert