



Industrie Service

The Testing and Calibration Laboratory
of TÜV SÜD Industrie Service GmbH

confirm to the company

Disruptive Technologies Research AS

Strandveien 17

NO-1366 Lysaker

because of the positive results of the tested temperature recorder,

Cloud Connector EU 4G with wireless sensors

according to the test-report No. **1132 TS-Rev. 1** from the 2022-02-24,
that the temperature recorder fulfils the requirements of the
standard DIN EN 12830:2018-10.

Sensor type	Industrial Gen 2	Standard Gen 2
Accuracy class	1	0,5
Measurement range	-40 °C to +55 °C	-40 °C to +55 °C
Application range	storage	
Software verification level	IIIb	

The manufacturer is allowed to mark the temperature recorder with the
following test-report number:

ATP-MUC 1132 TS

This certificate
is valid to: **2027-02-23**

(As long as the normative requirements, the declared fields of application, the product and the
production conditions are not significantly changed)

Munich, the 2022-02-24

Center of Competence
Refrigeration and air-conditioning
Head of test laboratory

Expert


Andreas Klotz


Thorsten Bennert